



Media Preparation Regulatory Demands – why?



The correct and repeatable media preparation has significant influence on the reproducibility of the dissolution test results

Following parameters are important:

- 1. precise and repeatable mixing of the media
- 2. warming of the media
- 3. degassing of the media
- 4. precise dispensing of the media into the vessels
- 5. documentation of the media preparation and dispensing

Media Preparation

Regulatory Demands – mixing, heating, dispensing



USP:

"Place the stated volume of the dissolution medium with an ...

- ... accuracy of 1% in the vessels ...
- ... and equilibrate it to 37 +/- 0,5 °C..."

Media Preparation

Regulatory Demands – mixing, heating, dispensing



Conclusion:

- 1. The "Stated Volumes" when mixing of all components of the media must be filled cold at "Standard Conditions"* into the vessels and then heated up.
- In accordance with the USP e.g. 900ml dispensed into a vessel weigh 897,4g. Only the weight of a liquid is not changing when warming it! Density and volume are changing!
- 3. It would be wrong to measure warm media in a volumetric flask or device calibrated at 20°C.
- 4. It would be wrong to use the specific weight (density) for gravimetric volume measuring of the media at e.g. 37°C.

* "Standard Conditions" for filling the vessels: USP → 25°C (USP 34), EP → 20°C and 1bar: 1 Litre of Water weighs 997,1g (USP) and 998,3g (EP)






USP:

"The USP recognizes that dissolved gases in the dissolution media may affect dissolution test results and ...

... recommends that gases be removed by filtering, warming and stirring under vacuum before the test is performed*."

FDA:

The FDA targeted the degassing more than ever as an area of concern. The recommended method* refers to parameters of the USP recommendation "filtering, warming, stirring, vacuum and process time". Stirring is realized by feeding air bubbles in the media bin. Additionally feeding air bubbles causes an Avalanche-Effect.



Media Preparation

Regulatory Demands – degassing



Experiences:

- there were not published many and clear experience reports on the consequences of various degassing methods till now *. Usually, internal company **experiences are not published**.
- a high theoretical degassing efficiency of any degasing method is reduced by fast **resaturation** of dissolved air in the media. The resaturation takes place during intermediate storage, during filling, during waiting periods and in the initial phase of the dissolution test.
- under dissolution test conditions with degased media at least the dissolved air in the media may not be oversaturated, i.e. **no formation of bubbles may arise**.

• In case of water

- with a temperature of **20°C** and an ambient air pressure of **1bar**, there are **9,08ppm** oxygen and air components dissolved.
- In case of **37°C** there are **6,72ppm**.
- If the temperature is increased to **100°C** and / or the air pressure is reduced against **0 bar** (full vacuum), no more air can be solved (**0,0ppm**).

Media Preparation

Regulatory Demands – degassing



Conclusion:


- The degassing is necessary at least to avoid oversaturation.
- The media can be dagassed either with the parameter "temperature" or "vacuum".
- •The results of different degassing methods must be compared after dosage into the vessel or at the start of the dissolution test.

Media Preparation Regulatory Demands – documentation



GLP / GMP:

The GLP / GMP rules demand a prescribed handling method SOP for the production of the media and the documentation of all working steps has to be done (traceability).





➤ easy handling:

The handling must be simple and convenient. Efficiency is increased only in connection with the...

... full acceptance of the equipment by the laboratory staff.

> space saving:

Nowhere in the labs is enough space available... ... such a unit must be compact.

\succ time and cost saving:


Media filtering, heating, mixing, and degassing must be performed without user intervention in a reasonable time...

... for increasing throughput, avoiding manual media handling and waiting for warming the media before the start of the dissolution test. The compliance of the regulations and standards saves a lot of operational and administrative work.

> USP, EP, FDA, GLP/GMP comformity:

There are many different requirements from regulations and standards, generating a lot of additional work...

... such a unit has to comply all requirements and has to save all associated operational and administrative work.





The following demands are important for an automated media preparation system.

Reproducibility and regulatory demands:

- 1. precise and repeatable mixing of the media
- 2. warming of the media
- 3. degassing of the media
- 4. precise dispensing of the media into the vessels
- 5. documentation of the media preparation and dispensing.

Additional user demands:

- 6. easy operation / handling
- 7. space saving
- 8. time and cost saving with less manual handling and less manual documentation
- 9. conformity with USP, EP, FDA, GLP/GMP





1. Demand: Precise and repeatable minxing of the media

The Dissoprep X8 / X15 is the original media preparation system which is working according the **gravimetrical principle** since 1998.

> Precise:

- GRAVIMETRICAL PRINCIPLE (lockout spec)
 - the weight is always measured and controlled also for foam
 - no additional calibration necessary for each media
- NO FLUDIC PUMP (lockout spec)
 - independent of accuracy of pumps
 - no problems with foaming media

> Repeatable:

- always the same process of media preparation for each method
- no individual intervention possible by the user (e.g. no additional degassing or heating after finished media preparation possible)





1. Demand: Precise and repeatable minxing of the media

The Dissoprep X8 / X15 is filtering, mixing, warming, degassing and dispensing the media according the USP, FDA, EP, GLP / GMP.

\succ filtering:

by an *easily exchangeable Filter Cartridge*, the remaining filter capacity is checked automatically, when necessary the system prompts the user to change the filter. The Flow Through Principle prevents bacterial growth.

> Mixing:

For the precise addition of concentrated hydrochloric acid, buffer or surfactant (SLS) a **second inlet channel** is provided.

The composition of additive and water is *controlled gravimetricly* by a Precision Load Cell.

An *electronic stirrer* mixes the components in the storage tank and ensures a homogenous mixture concentration (functionally monitored).

The precision of the composition is very high (deviation lower than 0,2% typ.).

DissoPrep X8 / X15 features (3)



2. Demand: warming of the media

A *special continous-flow heater warms the media* before degassing. This generates an enhanced degassing and saves considerable time when heating in the dissolution tester.

This heating process also ensures, that the warming is done reproducible, because no additional heating is possible, after the media is prepared.





3. Demand: Degassing of the media

The media is exposed to a *high vacuum* during withdrawing the raw media.

The interaction of heating, mixing and degassing generates an *effective de-aeration of the medium e.g. for water with typically 3-5ppm O*₂ (after filling into the vessel).

The USP does not specify degassing in figures because the dissolved oxygen measurement is not robust and no traceable standards for the calibration of oxygen meters are available.

Only the physical effect paramaters temperature, vacuum and duration of media exposure can be measured reliably.

DissoPrep X8 / X15

3. Demand: Degassing of the media


Degassing Efficiency Investigation with DissoPrep X8 © 2014 www.riggtek.com Dipl-Ing. Hansjuergen Riggenmann


As long as the vacuum is <300mbar and the temperature is > 32°C, a sufficient degassing of typical < 5ppm will be reached The DissoPrep-Routine-Applications (vacuum <100mbar, temperature between 32°C and 37°C, additional degassing time of 120 seconds) are resulting in a fast and very good degassing result!

Oxymeter from WTW OXI 330 DissoPrep X8 with Firmware 8.01

features (5)


				-										-					
Method Parameter	Test No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18
Temperature set.	°C	20	32	37	42	20	32	37	42	20	32	37	42	20	32	37	42	37	37
Temperature eff.	°C/10	154	325	365	410	205	324	366	410	212	323	365	408	206	322	389	408	367	386
min. Pressure (vacuum)	mbar	489	490	484	428	269	295	274	268	197	190	193	197	98	99	99	100	99	98
Volume	mL	1x 5400	1x5400	1x 5400	1x 5400	1x5400	1x 5400												
add. Degas Time	Sec	120	120	120	120	120	120	120	120	120	120	120	120	120	120	120	120	240	480
start Oxygen	ppm O ₂	9,1	8,3	8,1	9,2	8,4	8,4	8,9	8,9	8,6	8,1	8,5	8,2	8,6	8,2	8,4	8,4	8,6	8,5
end Oxygen	ppm O ₂	8,2	6,7	5,9	5.7	6,5	4,7	4,7	3,8	6,1	3,7	3,4	3,3	4,6	3,1	2,8	2,0	2,7	2,7
DPX8 routine application)											YES	YES			





3. Demand: Degassing of the media






Not degassed



Degassed





4. Demand: precise dispensing of the media in the vessels

The vessels in the dissolution tester are filled directly with a Dispense Tube.

The *highly precise dosage (<1% at 500 - 8.000g / 15.000g, typ. 2g)* of 900mL needs approximately *25 seconds* and is controlled by the Precision Load Cell.

The target volume per vessel can be selected between 100g and 8.000g / 15.000g (weight equivalents of mL).

DissoPrep X8 / X15 features (8)




5. Demand: documentation

All internal dosage processes are accurately controlled and *monitored by a Precision Load Cell*.

The individual dosage processes are documented and can be printed on an **external** *printer (parallel, USB or LAN) or administrated via the (Web-)Browser of any* PC (Browser-Interface for DissoPrep BI-DPX – standard option).

After each DISPENSE cycle the DissoPrep provides a **DISPENSE Protocol** containing the weights, the mixing ratio, the vacuum and the temperature. Also a **CALIBRATION Protocol** is provided. **The DissoPrep can be calibrated trough a software guided procedure easily.**





MEDIA DISPENSE REPORT No: 1

DissoPrep X8 RIGGTEK GmbH Germany

Serial Number : R41020520 Firmware Version: 8.08

General Data: Nominal filter capacity [1]: 5000 Remaining filter capacity [1]: 4948 Volume throughput up to now : 52

Method: 0 1000, 10.0, 6, 37.0 Method 0

Result of the dosages [g]:

			MEDIUM	ADDTV	RATIO	DEV8
Fill No	omina	11:	6000	60.0	0.010	
Fill A	ctual	l:	6538	65.4	0.010	+0.0
			MEDIUM	DEV8	ADDTV	DEV&
Vessel	No.	6:	1000	-0.0	10.0	+0.0
Vessel	No.	5:	1001	+0.1	10.0	+0.0
Vessel	No.	4:	1000	+0.0	10.0	+0.0
Vessel	No.	3:	1000	+0.0	10.0	+0.0
Vessel	No.	2:	1000	+0.0	10.0	+0.0
Vessel	No.	1:	1000	+0.0	10.0	+0.0
Result	(ave	era	ge):			

1000 0.0 Max.deviation: 1 Std.deviation: 0.3

Temperature (average): 37.1 C

MAX.VACUUM at 89 mbar abs. pressure Overall DEGASSING TIME (mm:ss): 10:30

Date, Time: 19.05.2016, 08:01

Name:

Signature:

Dates/Times no verified specifications

Printed at: 2018.8.23 12:50:02 by reviewer Submited at: 2018.8.22 10:55:18 by usertest1 Reviewed at: 2018.8.23 12:38:16 by reviewer Unsigned at: 2018.8.23 12:44:26 by approver

5. Demand: documentation

All parameters of a method and the instrument details are printed on the report as well as the details of the electronic signature*.

DissoPrep X8 / X15 features (10)



MANUAL CALIBRATION PROTOCOL									
No: 7									
for the media	for the media dosage								
with DOSAPREP	X8								
Serial Number	: 201000	056							
Firmware Version: 4.25									
General Data:									
Nominal filt	cer capacit	cy [l]:	5000						
Remaining filt	cer capacit	cy [l]:	4950						
Volume through	nput up to	now :	50						
Result of the	quantity m	neasurings	[g]						
II	NTERN EX	XTERN	DEV%						
No. 8:	1003	1005	-0.2						
No. 7:	1004	1004	+0.0						
No. 6:	1004	1005	-0.1						
No. 5:	1004	1003	+0.1						
No. 4:	1003	1003	+0.0						
No. 3:	1004	1003	+0.1						
No. 2:	1004	1005	-0.1						
No. 1:	1003	1002	+0.1						
Result (averag	ge):								
	1004	1004	-0.0						
Result of the	temperatu	re measurir	ıgs[C]						
II	NTERN E2	XTERN	DEV						
No. 8:	36.2	36.2	+0.0						
No. 7:	36.3	36.3	+0.0						
No. 6:	36.2	36.2	+0.0						
No. 5:	36.2	36.2	+0.0						
No. 4:	36.3	36.3	+0.0						
No. 3:	36.3	36.3	+0.0						
No. 2:	36.3	36.3	+0.0						
No. 1:	36.3	36.3	+0.0						
Result (averag	-								
	36.2	36.2	+0.0						
Date, Time:									
•••	• • • • • • • • • • •	• • • • • • • • • • •							
Name:									
	• • • • • • • • • • •	• • • • • • • • • • •							
Signature:									
	· · · · · · · · · · · · · · · · · · ·								
2000-10-03 2	ST·ST:ST ((JTC)							

5. Demand: documentation

Calibration details are reported separetely as well as performance test details.





5. Demand: documentation

Н	lome Method	Is Reports	Calibrations	Performance	Usermanager Rolemanage	r Setup Logout
Show 1	10 🗸 entries					Search:
#	 Status 	Report		Parameter	Date	Actions
1	Approved	MEDIA DISPENSE F	EPORT	Method: 0	2016-05-19 08:01:27	Print
2	Reviewed	MEDIA DISPENSE F	EPORT	Method: 0	2016-07-15 11:39:25	Print Sign
3	Unsigned	MEDIA DISPENSE F	EPORT	Method: 0	2016-07-15 12:11:36	Print Sign
4	Unsigned	MEDIA DISPENSE F	EPORT	Method: 0	2016-07-20 14:28:16	Print
5	Unsigned	MEDIA DISPENSE F	EPORT	Method: 10	2016-07-28 11:53:15	Print Sign
6	Unsigned	MEDIA DISPENSE F	EPORT	Method: 9	2016-09-05 09:40:21	Print Sign
7	Unsigned	MEDIA DISPENSE F	EPORT	Method: 0	2016-09-19 12:37:45	Print Sign
8	Unsigned	MEDIA DISPENSE F	EPORT	Method: 0	2016-09-22 13:10:59	Print Sign
9	Unsigned	MEDIA DISPENSE F	EPORT	Method: 0	2016-09-23 11:56:35	Print Sign
10	Unsigned	MEDIA DISPENSE F	EPORT	Method: 0	2016-07-15 13:13:35	Print Sign
Shov						3 4 5 10 N
	Home	Methods	Reports	Calibratio	ns Performance	

# 4	Status		
1	Submited	MANUAL PRESSURE CALIBRATION REPORT	
2	Unsigned	MANUAL TEMPERATURE CALIBRATION REPORT	
3	Unsigned	MANUAL TEMPERATURE CALIBRATION REPORT	Show 10 - entries
4	Unsigned	MANUAL QUANTITY CALIBRATION REPORT	# 🔺 Status 🔶 Report
5	Unsigned	MANUAL QUANTITY CALIBRATION REPORT	6 Reviewed MANUAL PERFORMANCE TEST REPORT
10	Unsigned	MANUAL PRESSURE CALIBRATION REPORT	7 Unsigned MANUAL PERFORMANCE TEST REPORT
11	Unsigned	MANUAL TEMPERATURE CALIBRATION REPORT	8 Unsigned MANUAL PERFORMANCE TEST REPORT
12	Unsigned	MANUAL TEMPERATURE CALIBRATION REPORT	9 Submited MANUAL PERFORMANCE TEST REPORT
13	Unsigned	MANUAL QUANTITY CALIBRATION REPORT	14 Unsigned MANUAL PERFORMANCE TEST REPORT
			21 Unsigned MANUAL PERFORMANCE TEST REPORT
15	Unsigned	MANUAL QUANTITY CALIBRATION REPORT	26 Unsigned MANUAL PERFORMANCE TEST REPORT

Showing 1 to 10 of 19 entries

Showing 1 to 7 of 7 entries





6. Demand: easy operation / handling

The standard Browser-Interface for DissoPrep *BI-DPX* allows to

- connect your *DissoPrep* via LAN to your local PC or to your company network
- and to administrate your *DissoPrep* easily via your (Internet-) Browser without any further software-installation and softwarevalidation, to
- administrate easily methods, reports incl. electronic signature* and 21 CFR compliant*,
- administrate user permissions by individual user roles,
- ensures data integrity,
- audittrail review,
- data backup*,
- parent-child-coupling,
- etc.



DissoPrep X8 / X15 features (13)



6. Demand: easy operation / handling

Beside the Browser-Interface, the user-interface of the DissoPrep is quite simple with only a few buttons for daily work.

The Remote Control Nozzle simplifies the direct dispensing into the vessels. The automated process of media preparation and dispensing – especially when using acids – minimizes *the risk to laboratory staff.*



7. Demand: space saving

The dimensions of the DissoPrep X8 / X15 are W 30cm x H 66cm x D 59cm quite compact to fit in every lab. Available as table or mobile device!





8. Demand: time and cost saving

The saving of time is considerable. E.g. the DissoPrep X8 prepares up to 8L of dissolution media automated in less than 15 minutes (12 minutes with prepared media). The fast and precise dosage by a tube into the individual vessels (~25 seconds for 900ml) saves timeconsuming handling steps. The preheated media saves up to 45 minutes for media heating with a conventional water bath dissolution tester.

If you calculate with:

30000

25000

20000

15000

10000

5000

0


n

- 180€ full costs of a manual test and
- 30€ costs of a test with the DissoPrep X8,

the ROI will already be reached after 90 tests



Profitability Analysis



DissoPrep X8 / X15 features (15)



9. Demand: USP, EP, FDA, GLP/GMP comformity

DissoPrep fulfills all requirements of the media preparation according to *the demanded rules of the USP and EP*, as well as according to the *recommendations of the FDA and of the GLP/GMP*.



The best proof: *The USP is using our DissoPrep!*



Media Preparation

Summary of meeting demands



to remember...

The correct and always repeatable media preparation has significant influence on the reproducibility of the dissolution test results

Reproducibility and regulatory demands:

- 1. precise and repeatable mixing of the media
- 2. warming of the media
- 3. degassing of the media
- 4. precise dispensing of the media into the vessels
- 5. documentation of the media preparation and dispensing.

Additional user demands:

- 6. easy operation / handling
- 7. space saving
- 8. time and cost saving with less manual handling and less manual documentation
- 9. conformity with USP, EP, FDA, GLP/GMP




with the DissoPrep:









DissoPrep X8 or X15

- X8 with 8 liter tank
 X15 with 15l tank (gross)
 same housing dimensions
- > same working principle



RC-Nozzle

 comfortable remote-controled dispensing with buttons at handle bar at tubing's end
 especially if 3m distance (tubings length) is used



Browser-Interface

- standard feature
- easy administration of methods, reports and users
- easy access with your (Internet-) Browser
- no software-installation or software-validation necessary

DissoPrep X8 / X15 options / accessories (2)



Pressure-Reducer

 for continous, pressureless contection to DI-water tap
 installation has to be done from house technicans of customer





LabCart / mobile use

- comfortable use of DissoPrep at different places
- place for different reservoirs (available at RIGGTEK), printer and vessel-rack
- UPS (uninterruptable power supply)
- second more slim version available soon!

Vessel-Rack

- safe transportation of vessels to dissolution tester
- ➢ isolation keeps vessels warm
- ➤ available for 6 or 8 vessels







printers

- for immediate protocol / certification printout
- suitable network- or USBprinters in DINA4- or bonpaper-size

measuring equipment

- all necessary measuring equipment for qualification of the DissoPrep is available at RIGGTEK
- special instruments with automatic communication to DissoPrep for automated qualification available





Thank you for your attention! Do you have any questions?

- Visit our webpage <u>www.riggtek.com</u> or
- give us a call *Tel.:* +49 89 2302469-0 or
- email us to support@riggtek.com

We are available! your RIGGTEK-Team

